



School of Chemical Sciences

Mahatma Gandhi University

Priyadarshini Hills P. O, Kottayam, 686560, Kerala, India

Phone: Off. 0481 2731036

Email : office.scs@gmail.com

No. SCS /PLAN/2018/Q1

28.12.2018

QUOTATION NOTICE

Sealed quotations in TRIPLICATE in are invited for the **Supply and Installation of Melting point apparatus(1no.)** for the use in School of Chemical Sciences.

Technical specification for Melting point apparatus

The melting point apparatus should work at 230V/ 50Hz with temperature range up to 300 °C. Facility to test more than one sample simultaneously must be present. It should have Digital selection and four digit display of temperature with 0.1°C display resolution and display hold facility. Temperature accuracy must be less than or equal to ± 2.5 °C at the maximum temperature attainable. Melting point tubes should be supplied along with the instrument.

Terms and Conditions

1. Only the name and a brief description/specification of the items is given. Further clarifications, if any may be obtained from the undersigned
2. Confirmation of quotation is decided on the basis of the quality and suitability of the item, extent of warranty and on the price considerations.
3. Taxes and freight charges, if any, should be shown in the quotation.
4. This being an educational institution, is eligible for concessional rate of 5% GST .
5. As per University rules payment can be made only direct to the firm after successful installation.
6. **Detailed Brochure** of the items should be enclosed with the quotations.
7. The quotation should be valid for At least 90 days.

The cover containing the quotation should be super scribed with “Quotation for **Melting point apparatus**” and send to the Director so as to reach the same on or before **5 PM on 17.12.2018**

Sd/-

Professor & Director