



MAHATMA GANDHI UNIVERSITY

SCHOOL OF BIOSCIENCES

Priyadarshini Hills P. O Kottayam – 686 560 Kerala, South India

Email : schoolofbiosciences@mgu.ac.in

Phone: (0481) 2731035

No.SBS/01/Rusa/Water Bath/2023

Date: 18.08.2023

QUOTATION NOTICE

Sealed competitive quotations are invited for the supply of **Water Bath** for the **School of Biosciences, Mahatma Gandhi University, Kottayam**. The quotation should clearly mention the make, model, specification and warranty of the items as well as the terms and conditions. The amount quoted should be inclusive of all taxes, excise duty as well as loading, unloading and transportation charges if any. Low quality and damage items will be rejected. The details of the items required are as mentioned below:-

Details of items required	Quantity
<p style="text-align: center;"><u>WATER BATH</u></p> <ul style="list-style-type: none">• The Water Bath should have a PID temperature controller with dual line display and soft key.• The Water Bath should have a capacity (effective volume) of 14 Litres more• The Water Bath should have inner dimension of 320 x 300 x 150 mm.• The Water Bath should have a temperature range of Ambient +5 °C to 99.9 °C.• The Water Bath should be a full SS304 – GMP model.• It should have a leak free inner seamless die pressed chamber.• The chamber should have a round corners that enables easy chamber cleaning and no microbes harboured.• It should have a clear bottom with externally cladded heater.• The Water Bath should have a dry run safety alarm and floating overshoot alarm.• The Water Bath should have a control accuracy of ± 0.3 at 37 °C.• The Water Bath should have a uniformity of ± 0.3 at 37 °C.• The Water Bath should have a high resolution of 0.1 °C.• The Water Bath should have a Illuminated Rocker	01

Standards.

- It should have CE and ISO certificate.
- Electric Supply – 230 VAC, 50 Hz, Single phase.
- It should be Supplied with Test Report/ Calibration cert. traceable up to national standards.
- It should be Supplied with gabbled lid.

The cover containing the quotations should be super scribed “**Quotation for Water Bath**” and the same should reach the Head of Department, School of Biosciences, Mahatma Gandhi University on or before **7th September, 2023 at 2 pm**. The quotations will be opened at 3 Pm on the same day at the School of Biosciences, Mahatma Gandhi University.

P.D.Hills



HEAD OF THE DEPARTMENT

Head of Department
School of Bioscience

