



MAHATMA GANDHI UNIVERSITY
Priyadarshini Hills, Kottayam-686560

Quotation No: ACCESSD/MGU/RUSA MRP 7/2023-24/08

Date: 20.10. 2023

QUOTATION NOTICE

Sealed quotations in original including all taxes are invited for the supply of Pocket water quality tester. The cover containing the quotations should be superscripted with "*Quotation for the supply of **Pocket water quality tester** for Advanced Centre of Environment Studies and Sustainable Development*" and the same should reach the Principal Investigator, RUSA 2.0, Advanced Centre of Environment Studies and Sustainable Development, School of Environmental Sciences, Mahatma Gandhi University, Priyadarsini Hills P.O, Kottayam (Dist.), Kerala - 686 560 on or before 31.10.2023.

Sl. No.	Item	Specification
1.	Pocket water quality tester	Waterproof, Display type: LCD Min to Max pH: (-1 to 15), pH resolution: 0.01 Buffer recognition: USA 4.01, 7.00, 10.01 and NIST 4.01, 6.86, 9.18. Calibration: upto 3 points pH Resolution- 0.01/0.1 Min to Max conductivity: 0 μ S to 20 mS Min to Max TDS: 0 ppm to 10 ppt Min to Max Salinity: 0 ppt to 10 ppt Min to Max Temperature: 0 $^{\circ}$ C to 60 $^{\circ}$ C

PI
RUSA 2.0, MRP No. 7
ACCESSD