Dr. N. Radhakrishnan International Centre for Medical Innovation
MAHATMA GANDHI UNIVERSITY
Priyadarsini Hills, Kottayam, Kerala, India -686560

Dr. Radhakrishnan E. K.
Director (Hon.)

NRICMI/DHR-HRD Project/02/Tender/2026

Dated: 21.01.2026

SHORT-TENDER NOTICE

Principal Investigator, DHR- (HRD Scheme) Project, Mahatma Gandhi University, Kottayam, 686560invites sealed competitive Tenders from reputed firms for the purchase of the "Biosafety Cabinet" as per the specifications mentioned below. The sealed covers containing Tender details are to be superscribed "Tender for Biosafety Cabinet" and sent to The Director, Dr N Radhakrishnan International Centre for Medical Innovation (NRICMI), Room no 703, 6th Floor, Convergence Academia Complex, Mahatma Gandhi University, P.D Hills P.O, Kottayam- 686560. Submit two separate sealed covers for the two-bid tender, with one cover containing the financial bid and the other containing the technical bid.

Last Date of Receiving Tenders: 29.01.2026, Time: 3.00 pm

Opening date of Tenders: 30.01.26, Time: 10 am

Billing address:
Principal Investigator,
DHR- (HRD Scheme) Project,
Mahatma Gandhi University,
Kottayam, 686560

SPECIFICATIONS:

Biosafety Cabinet

Bio-Safety Cabinet should be Class II Type A2, 3 Feet.

The equipment should meet the Cleanliness class of ISO Standard 14644 -1 ISO Class 3 or better

The biosafety should have air filters of the following specifications or better:

e-mail: nricmi@mgu.ac.in Mob: +919847901149

Scanned with OKEN Scanner

a) ULPA Final = 0.12μ with an efficiency of 99.9%

b) ULPA Exhaust = 0.12μ with an efficiency of 99.9%

The Equipment should have a One-piece full view frameless sliding toughened glass door 6 mm thick with a Motorized door or better specifications.

The Internal Work Area should be Complete Stainless Steel with rounded joint less corners or better.

Single-piece work tray to contain spillage with sloped easy-to-swipe perimeter.

The Equipment should be ergonomically designed with angled cabinet front which is sloped at around 10° for enhanced comfort and reduced operator fatigue.

The Biosafety should have 2 Nos of tube light 1.5 feet length 15 Watt ultra violet lighting or better.

Note:

All the accessories and spares for installation and demonstration should be supplied by the firm along with the instrument. The instrument should meet all the specifications given in the tender and minor variations will be examined by the technical committee before acceptance/rejection of the tender. Loading, unloading and transportation costs will be met by the bidder. The instrument should be supplied in the Research lab, Room 703, Dr N Radhakrishnan International Centre for Medical Innovation (NRICMI), Mahatma Gandhi University, P.D Hills P.O, Kottayam- 686560.

For more details contact Dr N Radhakrishnan International Centre for Medical Innovation (NRICMI), Mahatma Gandhi University, P.D Hills P.O, Kottayam on all working hours.

Terms & Conditions:

1. Tender form can be downloaded from the official website of Mahatma Gandhi University (https://www.mgu.ac.in/uploads/2020/08/Tender-Form.pdf?x99264) and the Tender form rate-0.2% of the quoted amount (Lowest amount = Rs. 400/- +GST, Maximum amount = Rs.1500/- +GST) can be paid by online in the University online payment system and challan should be attached along with the Tender.

2. The Tender Amount should be including all taxes.

3. Tenure of Tender should be three months.

4. A preliminary agreement in document paper of Rs. 200/- has to be submitted along with the tender.

5. Ernest Money Deposit (EMD) calculated @ 1% of (quoted amount) the Purchase assessment cost (PAC) should be paid by through the University's online payment

Scanned with OKEN Scanner

(www.mgu.ac.in.online payment - miscellaneous) or Demand Draft in favour of The H O D, School of Biosciences and submitted along with the sealed tender.

6. An agreement must be submitted by the qualified bidder and should submit 5% of the PAC as the security deposit.

7. Tenders received after the last date and time should be rejected. The Registrar has the right to accept or reject the tender without any reason. If the tender opening day is a holiday it will open in the next office working day.

8. If not received the minimum number of tenders, the last date will be extended to the next 15 days.

9. If more tenders are received during the extended period, the first received tenders are also to be considered.

10. Final payment should be released only after the proper installation and satisfactory report from the Technical Officer, School of Biosciences.



Principal Investigator,
DHR- (HRD Scheme) Project,
Mahatma Gandhi University,
Kottayam, 686560

Scanned with OKEN Scanner